Box Size: 3.73 x 8.5 Overall Size: 27 x 37 Folds to: 3.375 x 7.4

COVIDIEN

GIA™

Cartridge with Tri-Staple™ Technology

PT00195507

(92)PT00195507

1

	A	B	C	D	E	F
	GIA80MTC	80 mm	PURPLE	4.0 mm 3.5 mm 3.0 mm	1.2 - 1.75 mm	1.5 - 2.25 mm
	GIA80XTC	80 mm	BLACK	5.0 mm 4.5 mm 4.0 mm	1.7 - 2.25 mm	2.25 - 3.0 mm
	GIA80TMC	80 mm	TAN	3.0 mm 2.7 mm 2.4 mm	0.75 - 1.25 mm	0.88 - 1.8 mm
	GIA60MTC	60 mm	PURPLE	4.0 mm 3.5 mm 3.0 mm	1.2 - 1.75 mm	1.5 - 2.25 mm
	GIA60XTC	60 mm	BLACK	5.0 mm 4.5 mm 4.0 mm	1.7 - 2.25 mm	2.25 - 3.0 mm
	GIA60TMC	60 mm	TAN	3.0 mm 2.7 mm 2.4 mm	0.75 - 1.25 mm	0.88 - 1.8 mm

en

BEFORE USING PRODUCT, READ THE FOLLOWING INFORMATION THOROUGHLY.

IMPORTANT!

This booklet is designed to assist in using this product. It is not a reference to surgical techniques.

This device was designed, tested and manufactured for single-patient use only. Reuse or reprocessing of this device may lead to its failure and subsequent patient injury. Reprocessing and/or resterilization of this device may create the risk of contamination and patient infection. Do not reuse, reprocess or resterilize this device.

DESCRIPTION

The GIA™ stapler with Tri-Staple™ technology places two triple staggered rows of titanium staples and simultaneously cuts and divides tissue between these two triple rows. The GIA™ stapler and cartridges with Tri-Staple™ technology are currently available in 60 mm and 80 mm length.

The GIA™ stapler and cartridges with Tri-Staple™ technology are available in three staple sizes to accommodate various tissue thicknesses: thin/medium, medium/thick and extra thick.

Cartridges with thin/medium staple size (tan) deploy three height-progressive rows of 2.4 mm, 2.7 mm and 3.0 mm titanium staples. Cartridges with medium/thick staple size (purple) deploy three height-progressive rows of 3.0 mm, 3.5 mm and 4.0 mm titanium staples. Cartridges with extra thick staple size (black) deploy three height progressive rows of 4.0 mm, 4.5 mm and 5.0 mm titanium staples. Please refer to the "**SPECIFICATION CHART**" for availability of staple sizes and cartridge lengths.

Each GIA™ stapler with Tri-Staple™ technology may be reloaded up to 7 times for a total of 8 firings per instrument.

Qualitative and Quantitative Data	
*Maximum amount of titanium staples	0.2598 g

*Maximum amounts per reload when using the largest size (GIA80XT), g=grams

The target population is the general population of surgical candidates with weights greater than 10 kilograms. Clinical evidence supporting safety and performance has not been established in pediatric, pregnant women, or breast-feeding women populations for this device. Surgeons should consider patient factors before deciding if the device is suitable for use.

The product is to be used by medical professionals qualified in the transportation, preparation, and use of surgical devices. The GIA™ stapler is intended for use in a sterile operating room environment in surgical procedures where surgical stapling is indicated.

1 SPECIFICATION CHART

A) CARTRIDGE REORDER CODES
B) STAPLE LINE LENGTH
C) CARTRIDGE COLOR
D) OPEN STAPLE SIZE
E) CLOSED STAPLE SIZE
F) TISSUE COMPRESSION REQUIREMENT

Refer to the GIA™ stapler with Tri-Staple™ technology package insert for specific use instructions.

Please reference the GIA™ stapler with Tri-Staple™ technology package insert for specific information on staple dimensions, closure, and tissue compression requirements.

INDICATIONS/INTENDED PURPOSE

The GIA™ stapler with Tri-Staple™ technology and GIA™ cartridge with Tri-Staple™ technology have applications in abdominal and thoracic surgical procedures for resection, transection and creation of anastomosis.

The intended purpose of the device is transection and resection of tissue, and/or creation of anastomosis.

CLINICAL BENEFIT

Surgical staples provide mechanical support for tissues at surgical sites, where applied, for the duration of the healing process.

Summary of safety and clinical performance (SSCP)

The Summary of Safety and Clinical Performance (SSCP) for this device can be found at https://ec.europa.eu/tools/eudamed. Search for the SSCP using the manufacturer and device name and the following elements as applicable: device model, reference number, and catalog number.

CONTRAINDICATIONS

1. The instrument should not be used on tissue such as liver or spleen where compressibility is such that closure of the instrument would be destructive.
2. The instrument should not be used to staple tissues that are necrotic, friable, or have altered integrity, e.g., ischemic or edematous tissues.
3. Tissue thickness should be carefully evaluated before firing any stapler. Refer to "SPECIFICATION CHART" and the "Tissue Compression Requirement" section in the GIA™ stapler product insert.
4. "Tissue Compression Requirement" refers to the tissue compression requirement for each staple size. If the tissue cannot comfortably compress to the specified minimum requirement, or compresses to less than this requirement, the tissue may be too thick or too thin for the selected staple size.
5. Do not use this device for vascular resection and transection.

WARNINGS AND PRECAUTIONS

1. Intended users are healthcare professionals who have been trained in applicable surgical procedures and approaches involving stapling devices prior to employing this device.
2. As with all surgical staplers, surgeons should consider specific patient factors before deciding if the device is suitable for use. For example, preoperative radiotherapy may result in changes to tissue. These changes may, for example, cause the tissue thickness to exceed the indicated range for the selected staple size. Careful consideration should be given to any pre-surgical treatment the patient may have undergone and in corresponding selection of staple size.
3. The instrument should not be used on the central circulatory system or the central nervous system. Avoid use of the instrument on the aorta.
4. Visually inspect prior to firing for inclusion of unintended anatomic structures within the staple line.
5. When positioning the stapler on the application site, ensure that no unintentional obstructions, such as clips, are incorporated between the anvil and the cartridge prior to clamping. Firing over an obstruction may result in incomplete cutting action and/or improperly formed staples.
6. Clamping and unclamping of delicate structures may result in damage to tissue irrespective of stapler firing.
7. The GIA™ stapler with Tri-Staple™ technology is designed to be fully disposable and should not be fired more than 8 times per instrument. Additional applications may cause improperly formed staples which may result in leakage or disruption of the staple line.
8. When using the instrument more than once during the same procedure, be sure that the anvil is clear of tissue, blood and staples following each application.
9. When used in the abdomen, make sure no omental or mesenteric vessels are caught between the cartridge and anvil assembly prior to locking the instrument.
10. The GIA™ stapler with Tri-Staple™ technology is provided STERILE and is intended for multiple use during a SINGLE procedure. DISCARD AFTER USE. Reuse, reprocessing and/or resterilization of this device may create the risk of contamination, infection, permanent impairment or life-threatening injury. The device is sterilized by ethylene oxide gas, which is identified by the Environmental Protection Agency as a known carcinogen.
11. The cartridges are provided STERILE and are intended for use in a SINGLE procedure only. DISCARD AFTER USE. Reuse, reprocessing and/or resterilization of this device may create the risk of contamination, patient infection, permanent impairment or life-threatening injury. Do not reuse, reprocess or resterilize this device. The device is sterilized by ethylene oxide gas, which is identified by the Environmental Protection Agency as a known carcinogen.
12. Ensure that the staples are compatible with the stapler. The GIA™ cartridges with Tri-Staple™ technology can only be used with the GIA™ staplers with Tri-Staple™ technology. Compatibility of other manufacturer's staples and staplers has not been determined.
13. Compatibility of other manufacturer's staple line reinforcement material for use with GIA™ cartridges with Tri-Staple™ technology has not been determined.
14. The instrument should not be used to staple tissue outside of the labeled limits for maximum and minimum tissue thickness.
15. Ensure to select a cartridge with the appropriate staple size for the tissue thickness. Overly thick or thin tissue may result in unacceptable staple formation.
16. Do not use the stapler with thin/medium staple size (tan) on any tissue that does not compress within the range of 0.88 mm – 1.8 mm in thickness.
17. Do not use the stapler with medium/thick staple size (purple) on any tissue that does not compress within the range of 1.5 mm – 2.25 mm in thickness.
18. Do not use the stapler with extra thick staple size (black) on any tissue that does not compress within the range of 2.25 mm – 3.0 mm in thickness.
19. All GIA™ staplers with Tri-Staple™ technology of the same length can be used for three color cartridges (tan, purple and black). There is no need to change the stapler if you are only replacing one color cartridge with another and not changing the length of cartridge.
20. After firing, always inspect the staple line for hemostasis.
21. Failure to advance the handle completely may result in incomplete staple formation, which may compromise the integrity of the staple line.
22. Do not rotate the firing knob during firing. Rotating the knob during firing may cause damage and possible instrument malfunction.
23. This device is provided STERILE for single use only.
24. The instrument and single use reload are provided STERILE and are intended for use in a SINGLE procedure only. Sterile unless packaging has been opened or damaged. Dispose of used devices and unused devices with opened or damaged packaging in accordance with the end-user's medical and biological waste disposal requirements. "Sharps" should be disposed of in accordance with end-user's sharps containment, handling, and disposal requirements. Use caution during handling and disposal to avoid injury.
25. A thorough understanding of the principles involved in laser and electrosurgical procedures is essential to avoid shock and burn hazards to both patient and operator(s), and damage to the instrument.
26. The device contains the following substance(s) define as CMR 1B in a concentration above 0.1% weight by weight:
   CO: Cobalt, CAS No. 7440-48-4; EC No. 231-158-0
27. The device contains nickel, use of this device could result in sensitization or an allergic reaction.

MRI SAFETY INFORMATION

A person with implanted staples from the Covidien GIA™ stapler with Tri-Staple™ technology may be safely scanned under the following conditions. Failure to follow these conditions may result in injury.

- Static Magnetic Field Strength (B_0): 1.5T or 3.0T
- Maximum Spatial Field Gradient: 40 T/m (4,000 gauss/cm)
- RF Excitation: Circularly Polarized (CP)
- RF Transmit Coil Type: There are no Transmit Coil restrictions
- RF Receive Coil Type: Any
- Operating Mode: Normal Operating Mode
- Maximum Whole-Body SAR: 2 W/kg (Normal Operating Mode)
- Maximum Head SAR: 3.2 W/kg (Normal Operating Mode)
- Scan Duration: 2 W/kg whole-body average SAR for 60 minutes of continuous RF (a sequence or back-to-back series/scan without breaks)
- MR Image Artifact: The presence of this implant may produce an image artifact.

ADVERSE EVENTS

While every attempt has been made to reduce patient and user risks, all surgeries using this device carry some residual risk, even when used by trained physicians. The potential adverse reactions associated with the use of the GIA™ stapler and cartridges with Tri-Staple™ technology are: anastomotic leak, infection that may lead to death, organ perforation, inflammation, pneumothorax (when used for thoracic indication), bleeding that may lead to death, pain, fistula, tissue damage, obstruction, nerve entrapment, ischemia, allergic reaction and stenosis/stricture.

In the event that a serious incident has occurred related to device use, immediately report the event to Covidien, the competent authorities, and any other regulators as required.

STORE AT ROOM TEMPERATURE.

fr

Cartouche avec technologie Tri-Staple™

de

Magazin mit Tri-Staple™ Technologie

• Typ VF prijímacej cievky: akýkoľvek
• Prevádzkový režim: normálny prevádzkový režim
• Maximálna celotelová SAR: 2 W/kg (normálny prevádzkový režim)
• Maximálna hlavová SAR: 3,2 W/kg (normálny prevádzkový režim)
• Trvanie skenovania: 2 W/kg celotelová priemerná SAR v trvaní 60 minút nepretržovaného RF (sekvencia alebo za sebou idúce série/skenovania bez prestávok)
• MR obrazový artefakt: prítomnosť tohto implantátu môže vytvoriť obrazový artefakt.

NEŽIADUCE UDALOSTI

SKLADUJTE PRI IZBOVEJ TEPLOTE.

no

Magasin med Tri-Staple™-teknologi

LES FØLGENDE INFORMASJON NØYE FØR PRODUKTET TAS I BRUK.

VIKTIG!

Dette heftet er laget for å være til hjelp ved bruk av produktet. Det er ikke en referanse for kirurgiske teknikker.

Denne enheten er utformet, testet og fremstilt kun for bruk på én enkelt pasient. Hvis denne enheten brukes flere ganger eller rengjøres for gjenbruk, kan det føre til at funksjonsfeil og dermed til skade på pasienten. Klargjøring for gjenbruk og/eller resterilisering av denne enheten kan medføre risiko for kontaminering og infeksjon hos pasienten. Enheten må ikke brukes flere ganger, klargjøres for gjenbruk eller resteriliseres.

BESKRIVELSE

GIA™-stiftemaskin med Tri-Staple™-teknologi setter to tredoble, forskjøvne rader titanstifter og kutter og deler samtidig vevet mellom de to tredoble radene. GIA™-stiftemaskin og magasiner med Tri-Staple™-teknologi er på nåværende tidspunkt tilgjengelig i lengder på 60 mm og 80 mm.

GIA™-stiftemaskin og magasiner med Tri-Staple™-teknologi er på nåværende tidspunkt tilgjengelig i tre stiftestørrelser for å passe til forskjellige vevstykkelser: tynn/medium, medium/tykk og ekstra tykk.

Magasiner med tynn/medium stiftestørrelse (brunt) bruker tre høydeprogressive rader med titanstifter på 2,4 mm, 2,7 mm og 3,0 mm. Magasiner med medium/tykk stiftestørrelse (lilla) bruker tre høydeprogressive rader med titanstifter på 3,0 mm, 3,5 mm og 4,0 mm. Magasiner med ekstra tykk stiftestørrelse (svart) bruker tre høydeprogressive rader med titanstifter på 4,0 mm, 4,5 mm og 5,0 mm. Se tilgjengelige stiftestørrelser og magasinlengder i «**OVERSIKT OVER SPESIFIKASJONER**».

Hver GIA™-stiftemaskin med Tri-Staple™-teknologi skal kun lades på nytt opptil 7 ganger, noe som gir totalt 8 avfyringer per instrument.

Kvalitative og kvantitative data	
*Maksimalt antall titanstifter	0,2598 g

*Maksimalt antall per omlastning ved bruk av største størrelse (GIA80XT), g=gram

Målpopulasjonen er den generelle populasjonen av kirurgiske kandidater som veier mer enn 10 kg. Kliniske bevis som støtter sikkerhet og ytelse, er ikke dokumentert hos pediatriske populasjoner eller populasjoner med gravide eller ammende kvinner for denne enheten. Kirurgen bør vurdere pasientfaktorer før de avgjør om enheten er egnet for bruk.

Produktet skal brukes av medisinsk personell som er kvalifisert for transport, klargjøring og bruk av kirurgiske enheter. GIA™-stiftemaskinen er beregnet på bruk i en steril operasjonssal i kirurgiske prosedyrer der kirurgisk stifting er indikert.

1 OVERSIKT OVER SPESIFIKASJONER

A) BESTILLINGSKODER FOR MAGASIN
B) LENGDE PÅ STIFTLINJE
C) STIFTEMAGASINFARGE
D) STØRRELSE PÅ ÅPEN STIFT
E) STØRRELSE PÅ LUKKET STIFT
F) KRAV TIL VEVSKOMPRESJON

Se pakningsvedlegget for GIA™-stiftemaskin med Tri-Staple™-teknologipakke for spesifikk bruksanvisning.

Pakningsvedlegget for GIA™-stiftemaskin med Tri-Staple™-teknologipakke inneholder detaljert informasjon om stiftstørrelse, stiftlukking og krav til vevskompresjon.

INDIKASJONER / TILTENKT FORMÅL

GIA™-stiftemaskin med Tri-Staple™-teknologi og GIA™-magasin med Tri-Staple™-teknologi kan brukes til reseksjon, transeksjon og anastomosedannelse ved abdominale og torakale kirurgiske inngrep.

Det beregnede formålet med enheten er transeksjon og reseksjon av vev og/eller opprettning av anastomose.

KLINISK FORDEL

Kirurgiske stifter gir mekanisk støtte for vev på operasjonsstedet, hvis de tas i bruk, gjennom hele tilhelingsprosessen.

Sammendrag av sikkerhet og klinisk ytelse (Summary of safety and clinical performance – SSCP)

Sammendrag av sikkerhet og klinisk ytelse (SSCP) for denne enheten finnes på https://ec.europa.eu/tools/eudamed. Søk etter SSCP ved å bruke navnet på produsenten og instrumentet og følgende elementer etter behov: instrumentmodell, referansenummer og katalognummer.

KONTRAINDIKASJONER

1. Instrumentet skal ikke brukes på vev som lever eller milt, som kan klemmes så mye sammen at det ødelegges når instrumentet lukkes.
2. Instrumentet skal ikke brukes til å stifte vev som er nekrotisk, skjørt eller har endret integritet, for eksempel iskemisk eller ødematøst vev.
3. Vevets tykkelse skal evalueres nøye før en stiftemaskin avfyres. Se «OVERSIKT OVER SPESIFIKASJONER» og avsnittet «Vevskompresjonskrav» i pakningsvedlegget til GIA™-stiftemaskinen.
4. «Krav til vevskompresjon» viser til krav til komprimering av vev for hver stiftestørrelse. Hvis vevet ikke kan komprimeres til eller det angitte minstekravet, eller komprimeres til mindre enn dette kravet, kan vevet være for tykt eller for tynt for den valgte stiftestørrelsen.
5. Ikke bruk denne enheten til vaskulær reseksjon og transeksjon.

ADVARSLER OG FORHOLDSREGLER

1. Beregnede brukere er helsepersonell som har fått opplæring i gjeldende kirurgiske prosedyrer og tilnærminger som involverer stiftenheten, før bruk av denne enheten.
2. Som med alle kirurgiske stiftemaskiner bør kirurger vurdere spesifikke pasientfaktorer før de avgjør om enheten er egnet for bruk. Preoperativ strålebehandling kan for eksempel føre til endringer i vev. Disse forandringene kan for eksempel føre til at vevstykkelsen overstiger indikert område for den valgte stiftestørrelsen. Det må tas nøye hensyn til eventuelle preoperative behandlinger pasienten kan ha gjennomgått, og stiftstørrelse skal velges i samsvar med dette.
3. Instrumentet skal ikke brukes på det sentrale sirkulasjonssystemet eller sentralnervesystemet. Unngå å bruke instrumentet på aorta.
4. Inspiser visuelt for avfyring for inkludering av utilsiktede anatomiske strukturer innenfor stiftelinjen.
5. Når du plasserer stiftemaskinen på applikasjonsområdet, må du sikre at ingen utilsiktede hindringer, slik som klips, er innlemmet mellom amboltten og magasinet før du klemmer. Avfyring over en obstruksjon kan føre til ufullstendig kutting og/eller misformede stifter.
6. Klemming og løsning av klemme på skjøre strukturer kan føre til skade på vev uavhengig av fyring av stiftemaskinen.
7. GIA™-stiftemaskin med Tri-Staple™-teknologi er beregnet på å bli kassert etter maksimalt 8 avfyringer per instrument. Ytterligere bruk kan forårsake deformerte stifter som kan resultere i lekkasje eller ruptur av stiftelinjen.
8. Kontroller at det ikke er vev, blod eller stifter på amboltten etter hver avfyring hvis instrumentet skal brukes flere ganger under samme inngrep.

sl

Vložek s tehnologijo Tri-Staple™

PRED UPORABO IZDELKA NATANČNO PREBERITE NASLEDNJE INFORMACIJE.

Kvalitativni in kvantitativni podatki	
*Največja količina titanovih sponk	0,2598 g

bg

Касета с технология Tri-Staple™

ПРЕДИ ДА ИЗПОЛЗВАТЕ ТОЗИ ПРОДУКТ, ПРОЧЕТЕТЕ ВНИМАТЕЛНО СЛЕДНАТА ИНФОРМАЦИЯ.

Качествени и количествени данни	
*Максимално количество титанови скоби	0,2598 г

ro

Cartuș cu tehnologie Tri-Staple™

CITIȚI CU ATENȚIE INFORMAȚIILE DE MAI JOS ÎNAINTE DE A UTILIZA PRODUSUL.

Date calitative și cantitative	
*Cantitatea maximă de capse din titan	0,2598 g

A SE PĂSTRA LA TEMPERATURA CAMEREI.

et

Tehnoloogial Tri-Staple™ töötav kassett

ENNE TOOTE KASUTAMIST LUGEGE TÄHELEPANELIKULT LÄBI JÄRGMINE TEAVE.

Kvalitatiivsed ja kvantitatiivsed andmed	
* Titaanklambrite maksimaalne kogus	0,2598 g

HOIDA TOATEMPERATUURIL.

lv

Kasetne ar Tri-Staple™ tehnoloģiju

PIRMS IZSTRĀDĀJUMA LIETOŠANAS RŪPĪGI IZLASIET TĀLĀK SNIEGTO INFORMĀCIJU.

Kvalitatīvie un kvantitatīvie dati	
* Maksimālais titāna skavu daudzums	0,2598 g

GLABĀJIET ISTABAS TEMPERATŪRĀ.

sr

Kertridž sa tehnologijom Tri-Staple™

PRE UPOTREBE PROIZVODA DETALJNO PROČITAJTE SLEDEĆE INFORMACIJE.

Kvalitativni i kvantitativni podaci	
*Maksimalna količina titanijumskih kopči	0,2598 g

ČUVAJTE NA SOBNOJ TEMPERATURI.

lt

Kasetė su „Tri-Staple™“ technologija

PRIEŠ NAUDODAMI GAMINĮ, ATIDŽIAI PERSKAITYKITE TOLIAU PATEIKTĄ INFORMACIJĄ.

Kokybiniai ir kiekybiniai duomenys	
* Didžiausias titano kabių skaičius	0,2598 g

LAIKYKITE KAMBARIO TEMPERATŪROJE.

hr

Uložak s tehnologijom Tri-Staple™

PRIJE UPOTREBE OVOG PROIZVODA POZORNO PROČITAJTE SLJEDEĆE INFORMACIJE.

Kvalitativni i kvantitativni podaci	
*Maksimalna količina kopči od titanija	0,2598 g

SKLADIŠTITI NA SOBNOJ TEMPERATURI.

zh

采用 Tri-Staple™ 技术的 GIA™ 钉匣

使用产品前，请仔细阅读以下信息。

定性和定量数据	
*最大钛缝合钉用量	0.2598 g

zh-t

採用 Tri-Staple™ 技術之縫合釘匣

使用本產品之前，請先詳閱以下資訊。

定性與定量資料	
*鈦質縫合釘重量上限	0.2598 g

請儲存於室溫下。

ko

Tri-Staple™ 기술이 적용된 카트리지

제품을 사용하기 전에 다음 내용을 자세히 읽으십시오.

정성적 및 정량적 데이터	
*티타늄 스테이플의 최대량	0.2598g

실온 보관.

id

Kartrid dengan Teknologi Tri-Staple™

BACA SEMUA INFORMASI BERIKUT DENGAN SAKSAMA SEBELUM MENGGUNAKAN PRODUK.

Data Kualitatif dan Kuantitatif	
*Jumlah maksimum staples titanium	0,2598 g

SIMPAN PADA SUHU RUANGAN.

vi

Hộp ghim với công nghệ Tri-Staple™

TRƯỚC KHI SỬ DỤNG SẢN PHẨM, VUI LÒNG ĐỌC KỸ NHỮNG THÔNG TIN BÊN DƯỚI.

Dữ liệu định tính và định lượng	
*Số lượng ghim titan tối đa	0,2598 g

BẢO QUẢN Ở NHIỆT ĐỘ PHÒNG.

mk

Кертриџ со технологија Tri-Staple™

ВНИМАТЕЛНО ПРОЧИТАЈТЕ ГИ СЛЕДНИТЕ ИНФОРМАЦИИ ПРЕД ДА ГО УПОТРЕБИТЕ ПРОИЗВОДОТ.

Квалитативни и квантитативни податоци	
*Максимална количина на спојувалки од титаниум	0,2598 g

ЧУВАЈТЕ НА СОБНА ТЕМПЕРАТУРА.

en: Staple rows, staggered staple pattern

en: Open

en: Intended compressed tissue thickness

STERILE EO **en:** Sterilized using ethylene oxide

en: Single sterile barrier system

en: Single Use

Rx ONLY **en:** Caution: U.S. Federal law restricts this device to sale by or on the order of a physician.

en: Do not resterilize

en: Do not use if package is damaged

en: Consult instructions for use

en: Caution

MD **en:** Medical device

MR **en:** MR Conditional

en: Made from 100% recycled fibres. Minimum 35% post-consumer content.

7440-48-4 **en:** Contains hazardous substances

EC REP **en:** Authorized representative in the European Community

REF **en:** Catalog number

en: Manufacturer

en: Use-by Date

LOT **en:** Batch Code

en: Importer

en: Date of manufacture

en: Package quantity

CE 0123 **en:** CE mark

CE 0123

Technical manuals
www.medtronic.com/manuals

© 2024 Covidien.

Covidien llc, 15 Hampshire Street, Mansfield, MA 02048 USA.

EC REP Medtronic B.V., Earl Bakkenstraat 10, 6422 PJ Heerlen, The Netherlands.

www.covidien.com

+1 800 633 8766

+1 763 514 4000

COVIDIEN, COVIDIEN with logo, and the Covidien logo are trademarks of Covidien. ™* Third-party brands are trademarks of their respective owners. All other brands are trademarks of a Covidien company.

05/2024 Rev. 02